



# ● ContourGT Objectives Chart

Turret Mountable, Standard Objective Series (Parfocal with each other)

Objective (Magnification <sup>1</sup> )	2.5X		5X		5XL		10XBF		10X		20X		50X		115X	
<b>Working Distance (mm)</b>	3.5		6.7		9.4		10.6		7.4		3.7		3.4		0.6	
<b>Numerical Aperture</b>	0.07		0.12		0.13		0.25		0.3		0.4		0.55		0.8	
<b>Max Slope on Shiny Surfaces (deg)<sup>2</sup></b>	3		5.5		5.9		N/A		11.3		18.9		26.7		39.1	
<b>Max Slope on Rough Surfaces (deg)<sup>3</sup></b>	62		65		65		N/A		70		72		81		87	
<b>Optical Resolution (µm)<sup>4</sup></b>	3.8		2.2		2.1		1.1		0.9		0.7		0.5		0.33	
<b>Tallest Sample: ContourGT-X (mm)</b>	101.6		101.6		101.6		101.6		101.6		101.6		101.6		101.6	
<b>Tallest Sample: ContourGT-K (mm)</b>	95.3		95.3		95.3		95.3		95.3		95.3		95.3		95.3	
<b>Vertical Resolution (nm)<sup>5</sup></b>	<0.1		<0.1		<0.1		<0.1		<0.1		<0.1		<0.1		<0.1	
	<b>FOV (X by Y) (mm)</b>	<b>Spatial Sampling (µm)</b>	<b>FOV (X by Y) (mm)</b>	<b>Spatial Sampling (µm)</b>	<b>FOV (X by Y) (mm)</b>	<b>Spatial Sampling (µm)</b>	<b>FOV (X by Y) (mm)</b>	<b>Spatial Sampling (µm)</b>	<b>FOV (X by Y) (mm)</b>	<b>Spatial Sampling (µm)</b>	<b>FOV (X by Y) (mm)</b>	<b>Spatial Sampling (µm)</b>	<b>FOV (X by Y) (mm)</b>	<b>Spatial Sampling (µm)</b>	<b>FOV (X by Y) (mm)</b>	<b>Spatial Sampling (µm)</b>
<b>Standard Camera</b>																
0.55x zoom	4.6 x 3.5	7.2	2.3 x 1.7	3.6	2.3 x 1.7	3.6	1.2 x 0.9	1.8	1.2 x 0.9	1.8	0.6 x 0.4	0.9	0.23 x 0.17	0.36	0.10 x 0.08	0.16
0.75x zoom	3.4 x 2.5	5.3	1.7 x 1.3	2.6	1.7 x 1.3	2.6	0.8 x 0.6	1.3	0.8 x 0.6	1.3	0.4 x 0.3	0.7	0.17 x 0.13	0.26	0.07 x 0.06	0.11
1.0x zoom	2.5 x 1.9	4.0	1.3 x 1.0	2.0	1.3 x 1.0	2.0	0.6 x 0.5	1.0	0.6 x 0.5	1.0	0.3 x 0.2	0.5	0.13 x 0.10	0.20	0.06 x 0.04	0.09
1.5x zoom	1.7 x 1.3	2.6	0.8 x 0.6	1.3	0.8 x 0.6	1.3	0.4 x 0.3	0.7	0.4 x 0.3	0.7	0.2 x 0.2	0.3	0.08 x 0.06	0.13	0.04 x 0.03	0.06
2.0x zoom	1.3 x 1.0	2.0	0.6 x 0.5	1.0	0.6 x 0.5	1.0	0.3 x 0.2	0.5	0.3 x 0.2	0.5	0.2 x 0.1	0.2	0.06 x 0.05	0.10	0.03 x 0.02	0.04
<b>High-Res Camera (except K0 model)</b>																
0.55x zoom	6.4 x 4.8	4.7	3.2 x 2.4	2.3	3.2 x 2.4	2.3	1.6 x 1.2	1.2	1.6 x 1.2	1.2	0.8 x 0.6	0.6	0.32 x 0.24	0.23	0.14 x 0.10	0.14
0.75x zoom	4.7 x 3.5	3.4	2.4 x 1.8	1.7	2.4 x 1.8	1.7	1.2 x 0.9	0.9	1.2 x 0.9	0.9	0.6 x 0.4	0.4	0.24 x 0.18	0.17	0.10 x 0.07	0.10
1.0x zoom	3.5 x 2.6	2.6	1.8 x 1.3	1.3	1.8 x 1.3	1.3	0.9 x 0.7	0.6	0.9 x 0.7	0.6	0.4 x 0.3	0.3	0.18 x 0.13	0.13	0.08 x 0.06	0.06
1.5x zoom	2.4 x 1.8	1.7	1.2 x 0.9	0.9	1.2 x 0.9	0.9	0.6 x 0.4	0.4	0.6 x 0.4	0.4	0.3 x 0.2	0.2	0.12 x 0.09	0.09	0.05 x 0.04	0.04
2.0x zoom	1.8 x 1.3	1.3	0.9 x 0.7	0.6	0.9 x 0.7	0.6	0.4 x 0.3	0.3	0.4 x 0.3	0.3	0.2 x 0.2	0.2	0.09 x 0.07	0.06	0.04 x 0.03	0.03
<b>High-Res Camera (K0 model)</b>																
0.55x zoom	3.5 x 2.6	2.7	1.8 x 1.3	1.4	1.8 x 1.3	1.4	0.9 x 0.7	0.7	0.9 x 0.7	0.7	0.4 x 0.3	0.3	0.18 x 0.13	0.14	0.07 x 0.05	0.05
0.75x zoom	2.6 x 1.9	2.0	1.3 x 1.0	1.0	1.3 x 1.0	1.0	0.7 x 0.5	0.5	0.7 x 0.5	0.5	0.3 x 0.2	0.3	0.13 x 0.10	0.10	0.05 x 0.04	0.04
1.0x zoom	1.9 x 1.5	1.5	1.0 x 0.7	0.8	1.0 x 0.7	0.8	0.5 x 0.4	0.4	0.5 x 0.4	0.4	0.2 x 0.2	0.2	0.10 x 0.07	0.08	0.04 x 0.03	0.03
1.5x zoom	1.3 x 1.0	1.0	0.6 x 0.5	0.5	0.6 x 0.5	0.5	0.3 x 0.2	0.3	0.3 x 0.2	0.3	0.2 x 0.1	0.1	0.06 x 0.05	0.05	0.03 x 0.02	0.02
2.0x zoom	1.0 x 0.7	0.8	0.5 x 0.4	0.4	0.5 x 0.4	0.4	0.2 x 0.2	0.2	0.2 x 0.2	0.2	0.1 x 0.1	0.1	0.05 x 0.04	0.04	0.02 x 0.02	0.02

# ContourGT Objectives Chart

## Non-Turret Mountable Objectives

	LWD Objectives (Parfocal with each other)						Through Transmissive Media Objectives (Parfocal with each other)						Low Magnification Objectives					
	2X		5X		10X		2X		5X		10X		20X		1.0X		1.5X	
<b>Magnification<sup>1</sup></b>	2X		5X		10X		2X		5X		10X		20X		1.0X		1.5X	
<b>Working Distance (mm)</b>	22		22		22		8.0-9.8 <sup>6</sup>		8.0-9.8 <sup>6</sup>		8.0-9.8 <sup>6</sup>		8.0-9.8 <sup>6</sup>		2.5		9.6	
<b>Numerical Aperture</b>	0.055		0.14		0.17		0.055		0.14		0.25		0.28		0.04		0.14	
<b>Max Slope on Shiny Surfaces (deg)<sup>2</sup></b>	2.4		5.9		7.8		2.4		5.9		11.3		13		0.8		1.8	
<b>Optical Resolution (µm)<sup>4</sup></b>	4.9		1.9		1.6		4.9		1.9		1.1		1.0		6.7		6.5	
<b>Tallest Sample: ContourGT-X (mm)</b>	66.8		66.8		66.8		70.9		70.9		70.9		70.9		60.7		60.7	
<b>Tallest Sample: ContourGT-K (mm)</b>	60.5		60.5		60.5		64.5		64.5		64.5		64.5		54.4		54.4	
<b>Vertical Resolution (nm)<sup>5</sup></b>	<0.1		<0.1		<0.1		<0.1		<0.1		<0.1		<0.1		<0.1		<0.1	
	FOV (X by Y) (mm)	Spatial Sampling (µm)	FOV (X by Y) (mm)	Spatial Sampling (µm)	FOV (X by Y) (mm)	Spatial Sampling (µm)	FOV (X by Y) (mm)	Spatial Sampling (µm)	FOV (X by Y) (mm)	Spatial Sampling (µm)	FOV (X by Y) (mm)	Spatial Sampling (µm)	FOV (X by Y) (mm)	Spatial Sampling (µm)	FOV (X by Y) (mm)	Spatial Sampling (µm)	FOV (X by Y) (mm)	Spatial Sampling (µm)
<b>Standard Camera</b>																		
0.55x zoom	5.8 x 4.3	9.0	2.3 x 1.7	3.6	1.2 x 0.9	1.8	5.8 x 4.3	9.0	2.3 x 1.7	3.6	1.2 x 0.9	1.8	0.6 x 0.4	0.9	11.5 x 8.7	18.0	7.7 x 5.8	12.0
0.75x zoom	4.2 x 3.2	6.6	1.7 x 1.3	2.6	0.8 x 0.6	1.3	4.2 x 3.2	6.6	1.7 x 1.3	2.6	0.8 x 0.6	1.3	0.4 x 0.3	0.7	8.4 x 6.4	13.2	5.6 x 4.2	8.8
1.0x zoom	3.2 x 2.4	4.95	1.3 x 1.0	1.98	0.6 x 0.5	0.99	3.2 x 2.4	4.95	1.3 x 1.0	2.0	0.6 x 0.5	1.0	0.3 x 0.2	0.5	6.3 x 4.8	9.9	4.2 x 3.2	6.6
1.5x zoom	2.1 x 1.6	3.3	0.8 x 0.6	1.3	0.4 x 0.3	0.7	2.1 x 1.6	3.3	0.8 x 0.6	1.3	0.4 x 0.3	0.7	0.2 x 0.2	0.3	4.2 x 3.2	6.6	2.8 x 2.1	4.4
2.0x zoom	1.6 x 1.2	2.5	0.6 x 0.5	1.0	0.3 x 0.2	0.5	1.6 x 1.2	2.5	0.6 x 0.5	1.0	0.3 x 0.2	0.5	0.2 x 0.1	0.2	3.2 x 2.4	5.0	2.1 x 1.6	3.3
<b>High-Res Camera (except K0 model)</b>																		
0.55x zoom	8.0 x 6.0	5.9	3.2 x 2.4	2.3	1.6 x 1.2	1.2	8.0 x 6.0	5.9	3.2 x 2.4	2.3	1.6 x 1.2	2.3	0.8 x 0.6	1.2	16.0 x 12.0 <sup>7</sup>	11.7	10.7 x 8.0	7.8
0.75x zoom	5.9 x 4.4	4.3	2.4 x 1.8	1.7	1.2 x 0.9	0.9	5.9 x 4.4	4.3	2.4 x 1.8	1.7	1.2 x 0.9	1.7	0.6 x 0.4	0.9	11.8 x 8.8	8.6	7.8 x 5.9	5.7
1.0x zoom	4.4 x 3.3	3.23	1.8 x 1.3	1.3	0.9 x 0.7	0.6	4.4 x 3.3	3.23	1.8 x 1.3	1.3	0.9 x 0.7	1.3	0.4 x 0.3	0.6	8.82 x 6.6	6.45	5.9 x 4.4	4.3
1.5x zoom	2.9 x 2.2	2.2	1.2 x 0.9	0.9	0.6 x 0.4	0.4	2.9 x 2.2	2.2	1.2 x 0.9	0.9	0.6 x 0.4	0.9	0.3 x 0.2	0.4	5.9 x 4.4	4.3	3.9 x 2.9	2.9
2.0x zoom	2.2 x 1.7	1.6	0.9 x 0.7	0.6	0.4 x 0.3	0.3	2.2 x 1.7	1.6	0.9 x 0.7	0.6	0.4 x 0.3	0.6	0.2 x 0.2	0.3	4.4 x 3.3	3.2	2.9 x 2.2	2.2
<b>High-Res Camera (K0 model)</b>																		
0.55x zoom	4.4 x 3.3	3.4	1.8 x 1.3	1.4	0.9 x 0.7	0.7	4.4 x 3.3	3.4	1.8 x 1.3	1.4	0.9 x 0.7	0.7	0.4 x 0.3	0.3	8.8 x 6.6	6.8	5.9 x 4.4	4.5
0.75x zoom	3.2 x 2.4	2.5	1.3 x 1.0	1.0	0.7 x 0.5	0.5	3.2 x 2.4	2.5	1.3 x 1.0	1.0	0.7 x 0.5	0.5	0.3 x 0.2	0.3	6.5 x 4.8	5.0	4.3 x 3.2	3.3
1.0x zoom	2.4 x 1.8	1.9	1.0 x 0.7	0.8	0.5 x 0.4	0.4	2.4 x 1.8	1.9	1.0 x 0.7	0.8	0.5 x 0.4	0.4	0.2 x 0.2	0.2	4.9 x 3.6	3.75	3.2 x 2.4	2.5
1.5x zoom	1.6 x 1.2	1.3	0.6 x 0.5	0.5	0.3 x 0.2	0.3	1.6 x 1.2	1.3	0.6 x 0.5	0.5	0.3 x 0.2	0.3	0.2 x 0.1	0.1	3.2 x 2.4	2.5	2.2 x 1.6	1.7
2.0x zoom	1.2 x 0.9	0.9	0.5 x 0.4	0.4	0.2 x 0.2	0.2	1.2 x 0.9	0.9	0.5 x 0.4	0.4	0.2 x 0.2	0.2	0.1 x 0.1	0.1	2.4 x 1.8	1.9	1.6 x 1.2	1.3

### Notes

1. Chart specifications are based on nominal magnifications. Actual magnification is calibrated to National Institute of Standards Technology (NIST) traceable calibration standards.
2. As measured on an optically smooth surface and 1X magnification selector lens.
3. As measured on a rough-polished Si wafer and 1X magnification selector lens.
4. Optical resolution based on Sparrow Criteria at 535nm.
5. As demonstrated by a PSI measurement on a SiC reference mirror.
6. Dependent on the index and thickness of the transmissive material.
7. The 1.0X objective and 0.55x zoom provide a maximum FOV of 16.5mm diameter.

### ● Bruker Nano Surfaces Division

Tucson, AZ • USA  
 Phone +1.520.741.1044/800.366.9956  
 productinfo@bruker-nano.com

[www.bruker.com](http://www.bruker.com)